Tender for Supply of Electrical Engineering lab Equipment" No. PAF: IAST-RLE-ME-ITB-112-21 opened on 16-05-2022

Lot #:1. Electric Machines

Eligibility:

Cosmos Science Corporation	Eligible	
Muslim Scientific Corporation	Eligible	
Telec Electronics and Machinery	Eligible	
Electrical Engineering Services	Not Eligible	Non-Blacklisting Certificate
-	Telec Electronics and Machinery	Telec Electronics and Machinery Eligible

Qualification:

S.No:	Name of Firm/bidder	Qualification	Remarks
1.	Cosmos Science Corporation	Qualified	
2.	Muslim Scientific Corporation	Qualified	
3.	Telec Electronics and Machinery	Qualified	

Technical Specification Criteria:

S.No:	o: Name of Firm/bidder Technical Specification	
1.	Cosmos Science Corporation	Meets technical specifications
2.	Muslim Scientific Corporation	Could not meet technical specification. DC motors must be greater than 1 Kw. Induction motor and synchronous motor current less than 3 A. Power Supply have no three- phase variable supply
3.	Telec Electronics and Machinery	Could not meet technical specification. Loads less than the required specs. Shunt Regulator specs not given. Universal motors less than required power ratings. Power Supply data not given. DC Machines excitation less than required specs.

Four (04) bids received in response to lot 1 Machine Lab Equipment were evaluated. The three bidders i.e. COSMOS Sciences Corporation, Muslim Scientific Corporation, and Telec Electronics & Machinery were found eligible according to basic eligibility criteria of Tender document. One bidder i.e. Electrical Engineering Services was found not meeting the basic eligibility criteria due to lack of certificate of Backlisting.

The three bidders i.e. COSMOS Sciences Corporation, Muslim Scientific Corporation, and Telec Electronics & Machinery meet the minimum qualification (70%).

Muslim Scientific Corporation, and Telec Electronics & Machinery did not meet the tendered specifications.

A single bidder (COSMOS Sciences Corporation) was found to meet the technical specifications as per tender document.

It is therefore recommended that the financial offers of the said bidder may be opened.

(Dr. Mahammad Aamir)

Assistant Professor,

Assistant Per Forson

Department at Harmonia Stary

Fale August Fachbackusheles.

Institute of Applied Sciencia & Isernatopy,
Haripar, Falsitan

Member Technical Evaluation Committee